


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANM-2-C-P-N-ED-EO*

S.N.: *4941*

Date: *MAY 15 89*



40 db ATTEN

10 VOLTS/DIV

PRF ≈ 12 MHz

- a) Output signal amplitude:
0 TO ± 15 VOLTS
- b) Pulse width:
0.2 TO 2.0 NSER
- c) Rise time:
≤ 100 PSEZ
- d) Fall time:
≤ 135 PSEZ
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V

50-60 Hz