


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PVL-2-C-PN*

S.N.: *4945*

Date: *APRIL 21 89*



*70 66 ATTEN*  
*≈ 320 V<sub>0.48</sub> / DIV*  
*PRF = 2.0 KHz*

- a) Output signal amplitude:  
*0 TO ± 350 V<sub>0.48</sub>*
- b) Pulse width:  
*5 TO 100 NSER*
- c) Rise time:  
*≤ 2 NSER*
- d) Fall time:  
*≤ 3 NSER*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-60 Hz*