


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-AV-EW-P*

S.N.: *4999*

Date: *JUNE 12 89*



- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
1 TO 50 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 1 NS
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:

40 dB ATTEN

~ 10 VOLTS/DIV

PRF = 50 KHz

(RISE TIME)

- g) Prime power:
OK
+15 ± 1.0 VOLT
350 mA (MAX)
[Signature]