


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVT-V-N-PS*

S.N.: *5002*

Date: *JULY 13 1989*



50 db ATTEN

x 32 VOLTS/DIV

PRF = 20 KHz

- a) Output signal amplitude:
0 TO -50 VOLTS
- b) Pulse width:
2 TO 100 NSER
- c) Rise time:
≤ 350 PSEZ
- d) Fall time:
≤ 2 NSER
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz