


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVI-V-HV3-C-M1*

S.N.: *5003*


Date: *AUG 17 89*



*A RANGE (RISE TIME)
60 dB ATTEN*



A RANGE (FALL TIME)



B RANGE (RISE TIME)

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
A) 4 TO 100 NS
B) 100 NS TO 1.0 US
- c) Rise time:
≤ 200 PS
- d) Fall time:
A) < 3 NS
B) < 4 NS
- e) PRF:
0 TO 5 MHz
- f) Jitter, stability:
- g) Prime power: *OK*

120/240V

50-60 MHz

[Signature]