


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVD-AV-2-C-W-P-TM*

S.N.: *5004*

Date: *AUG 1 1989*



Ⓐ *40 db ATTEN*  
*10 V/DIV*  
*PRF = 50 KHz*



Ⓑ *AS Ⓐ BUT*  
*1 nSEC/DIV*  
*(RISE TIME)*

- a) Output signal amplitude:  
*0 TO +10 VOLTS*
- b) Pulse width:  
*20 TO 500 nSEC*
- c) Rise time:  
*≤ 100 pSEC*
- d) Fall time:  
*≤ 135 pSEC*
- e) PRF:  
*0 TO 50 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*170/280 V*  
*50-60 Hz*