


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMP-2-C-EB-7*

S.N.: *5026*

Date: *JULY 18 1989*



*40 db ATTEN*

*10 DIV/DIV*

*PRF = 1.0 MHz*

a) Output signal amplitude:

*0 TO +10 VOLTS*

b) Pulse width:

*5 TO 100 NS*

c) Rise time:

*≤ 100 PSEC*

d) Fall time:

*≤ 135 PSEC*

e) PRF:

*0 TO 1.0 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*120 / 240 V*

*50 - 60 Hz*

*[Signature]*