


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-5-HV1-N

S.N.: 5029

Date: AUG 29 89



40 dB ATTEN

≈ 10 VOLTS/DIV

PRF = 500 kHz

- a) Output signal amplitude:
0 TO -40 VOLTS
- b) Pulse width:
 ≤ 350 PSEC
- c) Rise time: (FWHM)
- d) Fall time: NA
- e) PRF: 0 TO 1.0 MHz
- f) Jitter, stability: OK
- g) Prime power:
 $\pm 15 \pm 1.0$ VOLT

150 mA (max)