


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AG-3-N-24-M-C*

S.N.: *5030*

Date: *JULY 28 1989*



~ 60 db ATTEN
~ 1.00 VOLTS/DIV
PRF = 20 KHz

- a) Output signal amplitude:
> 500 VOLTS.
 - b) Pulse width:
≤ 5 nsec
 - c) Rise time: *(1770 F power)*
 - d) Fall time: *NA*
 - e) PRF: *NA*
 - f) Jitter, stability: *OK*
 - g) Prime power:
120/240 V
50-60 Hz.
- [Signature]*

orig. sent to customer