


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-4B-PS-N*
S.N.: *5031*
Date: *JUNE 22 89*



R_i = 50 Ω
PRF = 50 kHz

- a) Output signal amplitude:
0 TO -400 VOLTS
- b) Pulse width:
0.1 TO 100 µSEC
- c) Rise time:
≤ 20 NS/SEC
- d) Fall time:
≤ 20 NS/SEC
- e) PRF:
MIN PW: 0 TO 50 kHz
PW ≤ 5.0 µSEC: 0 TO 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50-60 kHz

[Signature]