


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-MV-HV3-C-M-PW*

S.N.: *5038*

Date: *JUNE 23 '89*



50 db ATTEN
32 VOLTS/DIV
PRF = 100 KHz

- a) Output signal amplitude:
0 TO ± 40 VOLTS
- b) Pulse width:
0.4 TO 2.0 NS
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 250 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 KHz

[Signature]