


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMP-AN-HV3-C-N-PN*  
S.N.: *5039*  
Date: *JUNE 23 89*



*50 dB ATTEN.*  
*≈ 32 VOLTS/DIV.*  
*PRF = 100 kHz*

- a) Output signal amplitude:  
*0 TO ± 40 VOLTS*
- b) Pulse width:  
*0.4 TO 2.0 NS*
- c) Rise time:  
*≤ 150 PSE*
- d) Fall time:  
*≤ 250 PSE*
- e) PRF:  
*0 TO 100 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50 - 60 Hz*

*[Signature]*