


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PR-33-EN-EA-P*
S.N.: *5045*
Date: *JUNE 9 89*



50 db ATTEN

≈ 32 VOLTS DIV

PRF = 20 KHz

- a) Output signal amplitude:
0 TO +100 VOLTS
 - b) Pulse width:
20 TO 200 NSER
 - c) Rise time:
≤ 5 NSER
 - d) Fall time:
≤ 5 NSER
 - e) PRF:
0 TO 20 KHz
 - f) Jitter, stability:
OK
 - g) Prime power:
+24 ± 1.0 VOLTS
300 mA (MAX)
- [Signature]*