


PULSE GENERATOR
PERFORMANCE CHECK

Model: PW6-1-P

S.N.: 5047

Date: JUL 19 1989



60 db ATTEN

100 VOLTS/DIV

PRF = 20 KHZ

- a) Output signal amplitude: 0 TO +250 VOLTS
- b) Pulse width: 54 NS
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 50 KHZ
- f) Jitter, stability: OK
- g) Prime power: 415 VOLT
250 mA (MAX)