



PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-4B-PW-C-05-CA-EU-T
S.N.: 5049
Date: Aug 15 1989



(A) PW NORM MAX
 $R_L = 50 \Omega$
PRF = 50 kHz



(B) Low PW
 $R_L = 50 \Omega$
PRF = 1 kHz

- a) Output signal amplitude:
0 TO +400 V TO
- b) Pulse width: 50 ns
- c) Rise time:
0.1 TO 100 ns
- d) Fall time:
 ≤ 20 ns
- e) PRF:
MIN PW: 0 TO 1 kHz
MAX PW: 0 TO 50 Hz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz