


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-A1-PW-CR1-P*

S.N.: *5055*

Date: *APR 21 '89*



$R_L = 50 \Omega$
PRF = 50 kHz

- a) Output signal amplitude:
0 TO +250 V
- b) Pulse width:
30 TO 50 NSEC
- c) Rise time:
 ≤ 10 NSEC
- d) Fall time:
 ≤ 10 NSEC
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 \pm 1.000V
250 mA (MAX)
[Signature]