


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMR-B4-C-P-05-EW-BPO-JM*

S.N.: *5056*

Date: *Aug 28 89*



60 db atten

100 VOLTS/DIV

PRF = 100 KHz

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
1 TO 5 NS
- c) Rise time:
≤ 0.4 NS
- d) Fall time:
≤ 0.4 NS
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V

50-60 Hz

[Signature]