


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-AN-EN-P*
 S.N.: *5064*
 Date: *JUNE 12 '89*



40 db ATTEN

~ 10 VOLTS/DIV
PRF = 50 KHz

- a) Output signal amplitude: *0 TO +15 VOLTS*
- b) Pulse width: *1 TO 50 NSER*
- c) Rise time: *≤ 100 PSER*
- d) Fall time: *≤ 1 NSER*
- e) PRF: *0 TO 50 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 ± 1.0 VOLTS*
350 mA (MAX)

[Signature]