


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-B3-EN-CA-P*
S.N.: *5065*
Date: *JUNE 9 89*



50 db ATTEN

∴ ≈ 32 VOLTS/DIV

PRF = 20 KHz

- a) Output signal amplitude:
0 TO +100.0 VOLTS
- b) Pulse width:
20 TO 200 NS
- c) Rise time:
≤ 5 NS
- d) Fall time:
≤ 5 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 ± 1.0 VOLTS
300 mA (MAX)

[Signature]