


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-P*
S.N.: *5068*
Date: *Aug 23 '89*



4046 ATTEN

5 VOLTS/DIV

PRF ≈ 7 MHz

- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.2 TO 6.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 ± 1.000V
400 mA
(MAX)

[Signature]