


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-HV1-C-P*

S.N.: *5071*

Date: *JUNE 1 '89*



*50 db ATTEN*  
*= 32 VOLTS/DIV*

*PRF = 50 KHz*

- a) Output signal amplitude: *+30 TO +100 V*
- b) Pulse width: *≤ 1.0 nSEC*
- c) Rise time: *C10% ASE (part)*
- d) Fall time: *NA*
- e) PRF: *0 TO 100 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*  
*50-60 Hz*