


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PA4-HV1-C-P*

S.N.: *5072*

Date: *JUNE 1 89*



50 db ATTEN

≈ 32 VOLTS / DIV

PRF = 100 KHz

- a) Output signal amplitude:
+30 TO +100 VOLTS
- b) Pulse width:
≤ 1.0 NSEC
- c) Rise time: *C10% RISE TIME*
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
- g) Prime power: *OK*
120/240 V
50-60 Hz.

[Signature]