


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-5-1-C-D*

S.N.: *5083*

Date: *JUNE 2 '89*



*40 db ATTEN*

*10.00V/DIV*

*PRF = 500 kHz*

- a) Output signal amplitude:  
*0 TO +10.00V*
- b) Pulse width:  
*≤ 130 psec*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-60 Hz*