


PULSE GENERATOR
PERFORMANCE CHECK


Model: *PT-4-171-CP*

S.N.: *5089*

Date: *JUNE 20 89*



*(A) 50db ATTEN
2 32 VOLTS/DIV*



*(B) AS @ BUT
500 PSEC/DIV
(RISE TIME)*

- a) Output signal amplitude:
0 TO +50 V
- b) Pulse width:
3 TO 100 NSEC
- c) Rise time:
≤ 250 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*