


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-MIP-P*

S.N.: *5090*

Date: *Aug 1 1989*



*50 dB ATTEN*  
*32 VOLTS/DIV*  
*PRF = 50 KHz*

- a) Output signal amplitude:  
*71 ± 40 VOLTS*
- b) Pulse width:  
*2 TO 100 NS*
- c) Rise time:  
*≤ 350 PSEC*
- d) Fall time:  
*≤ 2 NS*
- e) PRF:  
*MIN PW: 0 TO 1 MUTE*  
*MAX PW: 0 TO 20 KHz*
- f) Jitter, stability: *(SEE PAGES 1+2)*  
*OK*
- g) Prime power:

*+15 ± 1.0 VOLT*  
*150 mA (MAX)*