


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-5B-C-PN-CA-EN-M*


S.N.: *5092*

Date: *SEPT 18 1989*



- a) Output signal amplitude:
0 TO ±40 AMPS TO
- b) Pulse width: *0.75 ns*
25 TO 500 NSER
- c) Rise time:
≤ 10 NSER
- d) Fall time:
≤ 10 NSER
- e) PRF: *0 TO 1 KHz*
- f) Jitter, stability:
OK

Point $R_c = 0.7 \Omega$
 $PRF = 1.0 \text{ KHz}$



- g) Prime power:
120/240 V
50-60 Hz.

Point $R_c = 0.7 \Omega$
 $PRF = 1.0 \text{ KHz}$

[Signature]