


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-2-PN*

S.N.: *5093*

Date: *JULY 20 1989*



≈ 6 AMPS/DIV

PRF = 20 KHz

- a) Output signal amplitude:
0 TO ±10 AMPS
- b) Pulse width:
2 TO 15 NSER
- c) Rise time:
≤ 1 NSER
- d) Fall time:
≤ 2 NSER
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 VOLT
300 mA (MAX)

[Signature]