


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-5-1-PS-P*

S.N.: *5695*

Date: *04 27 89*



*40 dB ATTEN*

*10 VOLTS/DIV*

*PRF = 500 KHz*

a) Output signal amplitude:  
*0 TO 110 VOLTS*

b) Pulse width:  
*≤ 130 PSEC*

c) Rise time:  
*NA*

d) Fall time:  
*NA*

e) PRF:  
*0 TO 1 MHz*

f) Jitter, stability:

OK  
g) Prime power:  
*110/240 V*

*50-60 Hz*

*Q*