


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PI-MP-P*

S.N.: *5103*

Date: *JULY 31 1989*



*50 dB ATTEN*

*≈ 32 VOLTS/DIV*

*PUSE = 50 KHz*

a) Output signal amplitude:

*≥ +40 VOLTS*

b) Pulse width:

*2 TO 100 NS*

c) Rise time:

*≤ 350 PSEC*

d) Fall time:

*≤ 2 NS*

e) PRF:

*MIN PW : 0 TO 1 MHz*

*MAX PW : 0 TO 20 KHz*

f) Jitter, stability: *See Fig 1+2*

g) Prime power:

*OK*  
*+15 ± 1.0 VOLTS*

*150 mA MAX*

*AA*