


PULSE GENERATOR
PERFORMANCE CHECK

Model: PV6-1-SRIA-P

S.N.: 5117

Date: JUNE 27 89



60 dB ATTEN

~ 100 VOLTS/DIV

PRF = 50 KHz

a) Output signal amplitude:

$\approx +160$ VOLTS

b) Pulse width:

≤ 4 NSEC

c) Rise time:

≤ 200 PSEC

d) Fall time:

≤ 4 NSEC

e) PRF:

0 TO 50 KHz

f) Jitter, stability:

g) Prime power: ^{OK}

+15 \pm 1.0 VOLT

200 mA (MAX)