


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV6-1-SRIA-N

S.N.: 5118

Date: JUNE 27 1989



60 dB ATTEN

≈ 100 VOLTS/DIV

PRF ≈ 50 KHz

a) Output signal amplitude:

≈ -160 VOLTS

b) Pulse width:

≤ 4 NSEC

c) Rise time:

≤ 200 PSEC

d) Fall time:

≤ 4 NSEC

e) PRF:

0 TO 50 KHz

f) Jitter, stability:

g) Prime power: OK

$\pm 15 \pm 1.0$ VOLT

200 mA (MAX)