


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-M2-P*

S.N.: *5119*

Date: *AUG 2 1989*



*50 db ATTEN*

*32 V/DIV*

*PRF = 500 KHz*

a) Output signal amplitude:

*≥ 60 V/DIV*

b) Pulse width:

*≤ 2 NSEC*

c) Rise time:

*NA*

d) Fall time:

*NA*

e) PRF:

*0 TO 1.0 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+135 ± 5 V/DIV*

*5 mA (MAX)*

*[Signature]*