


PULSE GENERATOR
PERFORMANCE CHECK


Model: AVI-V-HV3-C-N

S.N.: 5124

Date: SEPT 19 89



Ⓐ 60 db ATTEN
100 VOLTS/DIV
PRF = 5.0 KHz



Ⓑ AS Ⓐ BUT
1.0 nSEC/DIV
(RISE TIME)

- a) Output signal amplitude:
0 TO -100 V.
 - b) Pulse width:
4 TO 100 nSEC
 - c) Rise time:
≤ 200 pSEC
 - d) Fall time:
≤ 3 nSEC
 - e) PRF:
0 TO 5 MHz
 - f) Jitter, stability:
OK
 - g) Prime power:
120/240 V
50-60 Hz
- [Signature]*