


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *APR-A-1PW-C-CA-UCI-P*

S.N.: *5126*

Date: *SEPT 7 1959*



*$R_i = 50 \Omega$*

*PRF = 5.0 KHz*

- a) Output signal amplitude:  
*0 TO +200 V.*
- b) Pulse width:  
*50 NSEC TO 100 NSEC*
- c) Rise time:  
 *$\leq 10$  NSEC*
- d) Fall time:  
 *$\leq 10$  NSEC*
- e) PRF:  
*PW MAX: 0 TO 5 KHz*  
*PW MIN: 0 TO 100 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*