


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-C-P*
S.N.: *5128*
Date: *SEPT 13 '89*



≈ 60 Amms/DIV
PRF = 1.0 kHz

- a) Output signal amplitude:
+25 TO +100 Amms
- b) Pulse width:
≤ 25 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz