


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PNH-1423-P-05A*

S.N.: *5129*

Date: *OCT 4 89*



*50 db ATTEN*  
*= 32 VOLTS/DIV*  
*PK.F = 700 KHZ*

- a) Output signal amplitude:  
*> + 60 VOLTS*
- b) Pulse width:  
*≤ 1 NSEC*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+ 15 ± 1.0 VOLT*  
*150 mA MITT*
- h)  $t_p$ : *≤ 20 NSEC*  
*(≈ 10 NSEC)*