


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV4-M43-P-05A*

S.N.: *5130*

Date: *Oct 4 89*



50 db ATTEN
i = 32 VOLTS/DIV
VRF = 700 kHz

- a) Output signal amplitude: *> +60 VOLTS*
- b) Pulse width: *≤ 1 nSEC*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *415 ± 1.0 VOLTS*
150 mA (MAX)
- h) t_p : *≤ 20 nSEC*
(= 10 nSEC)