


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-B3-C-P*

S.N.: *5137*

Date: *July 28 1989*



R_L = 50 Ω

PRF = 20 kHz

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
70 TO 200 NS
- c) Rise time:
5.5 NS
- d) Fall time:
5 NS
- e) PRF:
0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
30-60 Hz

[Signature]

orig sent to customer