


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-1-PW-PS-P*

S.N.: *5138*

Date: *JUN 7 31 1989*



R_L = 50 Ω
PRF = 1.0 kHz

- a) Output signal amplitude:
0 TO +200 V
- b) Pulse width:
0.1 TO 5.0 μ S
- c) Rise time:
 ≤ 10 nS
- d) Fall time:
 ≤ 10 nS
- e) PRF:
0 TO 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:

70/240 V
30-60 MHz
TD