


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVMP-2-C-P*

S.N.: *5141*

Date: *OCT 17 89*



Ⓐ 40 db ATTEN  
10 VOLTS/DIV  
PRF = 1.0 MHz



Ⓑ AS Ⓐ BUT  
1.0 NS/DIV

- a) Output signal amplitude:  
*0 TO +10 VOLTS*
- b) Pulse width:  
*5 TO 100 NS*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*

*A*