


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1-C-PN*

S.N.: *5146*

Date: *OCT 11 89*



40 dB ATTEN
± 10 VOLTS PN
PRF = 700 KHz

- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 35 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz
[Signature]