


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVI-V-C-P

S.N.: 5153

Date: OCT 10 89



50 dB ATTEN

$\approx 32$  VOLTS/DIV

PRF = 10 KHZ.

- a) Output signal amplitude:  
0 TO +50 VOLTS
- b) Pulse width:  
2 TO 100 NSER
- c) Rise time:  
 $\leq 350$  PSEER
- d) Fall time:  
 $\leq 2$  NSER
- e) PRF:  
0 TO 20 KHZ
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240 V  
50-60 HZ.

*[Signature]*