


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-3A-P-TM
S.N.: 5156
Date: AUG 22 89



$R_L = 5.0 \Omega$
60 db atten

$\approx 100 \text{ VOLTS/DIV}$

$\approx 20 \text{ AMPS/DIV}$

- a) Output signal amplitude:
+25 TO +75 Amp
- b) Pulse width:
2.5 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 \pm 1.0 V_{eff}
300 mA (max)