


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVL-2-e-PN*

S.N.: *5175*

Date: *SEPT 5 1989*



70 db ATTEN

± 320 VOLT/IN

PRF = 2.0 KHZ

- a) Output signal amplitude:
0 TO ± 350 V
- b) Pulse width:
5 TO 100 NSER
- c) Rise time:
≤ 2 NSER
- d) Fall time:
≤ 3 NSER
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz