


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVN-1-C-PN-M*

S.N.: *5204*

Date: *Nov 6 1989*



① PRF = 110 MHz
40 dB ATTEN
5 VOLTS/DIV



③ PRF = 260 MHz
40 dB ATTEN
5 VOLTS/DIV

- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0.2 TO 1.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 100 PSEC
- e) PRF:
50 TO 260 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz

[Signature]