


PULSE GENERATOR

PERFORMANCE CHECK

Model: *APR-3-pw-ps-USA1-P*

S.N.: *5205*

Date: *NOV 21 1989*



R_L = 5 K

PRF = 4 Hz

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
R_L > 5 K
0.5 TO 50 NSEC
- c) Rise time: (20% MAX DUTY CYCLE)
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC
- e) PRF:
MAX PW: 0 TO 4 Hz
MIN PW: 0 TO 400 Hz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz