


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-AV-HV3-CPN-OT-M*
S.N.: *5206*
Date: *OCT 23 1989*



50 dB ATTEN
32 V OUT/IN
PRF = 100 kHz

- a) Output signal amplitude:
0 TO ± 40 V OUT
- b) Pulse width:
0.4 TO 2.0 NS
- c) Rise time:
≤ 150 pSEC
- d) Fall time:
≤ 250 pSEC
- e) PRF:
0 TO 100 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

[Signature]