


PULSE GENERATOR
PERFORMANCE CHECK

Model: AGB-2 PN

S.N.: 5208

Date: SEPT 14 1989



$R_c = 5 \Omega$
 $\approx 6 \text{ AMPS/PN}$

$f_{RF} = 10 \text{ MHz}$

- a) Output signal amplitude: 0 TO ± 10 AMPS TO
- b) Pulse width: 5 nS
- c) Rise time: 2 TO 10 nS/R
- d) Fall time: 1 nS/R
- e) PRF: 6 2 nS/R
- f) Jitter, stability: 0 TO 20 MHz
- g) Prime power: OK

+15 \pm 1.0 V/ST.

300 mA (PN/A)