


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ACL-2-PS-N-EA-EW*

S.N.: *5214*

Date: *NOV 10 89*



*70 dB ATTEN*

*≈ 320 VOLTS/DIV*

*PRF = 2 KHz.*

- a) Output signal amplitude:  
*0 TO -350 VOLTS*
- b) Pulse width:  
*5 TO 100 NS*
- c) Rise time:  
*≤ 2 NS*
- d) Fall time:  
*≤ 3 NS*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-60 Hz*