


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVMMR-1-P-CA-EN-ED-TRF-ECL-VMI-JM*

S.N.: *5221*

Date: *OCT 30 89*



40 ohm ATTEN
5 VOLTS/DIV
7.0 MHz
PKF

- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
5 TO 50 NSEC.
- c) Rise time:
≤ 100 PSEC OR
≤ 1 NSEC
- d) Fall time:
≤ 135 PSEC OR
51 NSEC
- e) PRF:
0 TO 10 MHz
- f) Jitter, stability: *(20% MAX DUTY) CYCLE*
OK
- g) Prime power:

+24 ± 1.0 VOLTS
300 mA (MAX).

A